

**H₂O₂ 50 T****209****0.01 - 0.5 mg/l H₂O₂****DPD / Catalyst**

Instrument specific information

The test can be performed on the following devices. In addition, the required cuvette and the absorption range of the photometer are indicated.

Instrument Type	Cuvette	λ	Measuring Range
SpectroDirect, XD 7000, XD 7500	□ 50 mm	510 nm	0.01 - 0.5 mg/l H ₂ O ₂

Material

Required material (partly optional):

Reagents	Packaging Unit	Part Number
Hydrogen Peroxide LR	Tablet / 100	512380BT
Hydrogen Peroxide LR	Tablet / 250	512381BT

Application List

- Waste Water Treatment
- Drinking Water Treatment
- Raw Water Treatment
- Disinfection Control

Sampling

1. When preparing the sample, Hydrogen Peroxide outgassing, e.g. through the pipette or shaking, must be avoided.
2. The analysis must take place immediately after taking the sample.


Preparation

1. Cleaning of vials:
As many household cleaners (e.g. dishwasher detergent) contain reducing substances, this can lead to lower results with the determination of Chlorine. To avoid measurement errors, the glassware used should be free of chlorine consumption. To achieve this, all glassware should be placed in a sodium hypochlorite solution (0.1 g/l) for one hour and then rinsed thoroughly with deionised water.
2. The DPD colour development is carried out at a pH value of 6.2 to 6.5.
The reagents therefore contain a buffer for the pH adjustment. Strong alkaline or acidic water samples must therefore be adjusted between pH 6 and pH 7 before the analysis (use 0.5 mol/l Sulphuric acid or 1 mol/l Sodium hydroxide).


Implementation of the provision Hydrogen peroxide with Tablet

Select the method on the device


For this method, no ZERO measurements are to be carried out with the following devices: XD 7000, XD 7500




Fill 50 mm vial with sample.





Place **sample vial** in the sample chamber. • Pay attention to the positioning.



Press the **ZERO** button.




Remove **vial** from the sample chamber. Empty vial.




Dry the vial thoroughly.


For devices that require **no ZERO measurement**, start here.




Rinse a beaker **with the sample** and empty it, leaving a few drops remaining in the beaker.




Add **HYDROGENPEROXIDE LR tablet**.




Crush tablet(s) by rotating slightly.




Put **10 ml sample** in the sample vessel.




Dissolve tablet(s) by inverting.




Fill **50 mm vial** with **sample**.



Place **sample vial** in the sample chamber. • Pay attention to the positioning.



Press the **TEST** (XD: **START**) button.



Wait for **2 minute(s) reaction time**.

Once the reaction period is finished, the measurement takes place automatically.

The result in mg/l Hydrogen Peroxide appears on the display.

Chemical Method

DPD / Catalyst

Appendix

Interferences

Persistent Interferences

1. All oxidising agents in the samples react like hydrogen peroxide, which leads to higher results.

Removeable Interferences

1. Concentrations above 5 mg/l hydrogen peroxide can lead to results within the measuring range of up to 0 mg/l. In this case, the water sample must be diluted with water that is free from hydrogen peroxide. 10 ml of the diluted sample should be mixed with the reagent and the measurement taken again (plausibility test).

Bibliography

Colorimetric Chemical Analytical Methods, 9th Edition, Lovibond

Derived from

US EPA 330.5

APHA 4500 Cl-G

^{a)} determination of free, combined and total | ^{b)} Reactor is necessary for COD (150 °C), TOC (120 °C) and total -chromium, - phosphate, -nitrogen, (100 °C) | ^{c)} MultiDirect: Adapter is necessary for Vacu-vials® (Order code 19 20 75) | ^{d)} Spectroquant® is a Merck KGaA Trademark | ^{e)} alternative reagent, used instead of DPD No.1/No.3 in case of turbidity in the water sample caused by high concentration of calcium and/or high conductivity | ^{f)} additionally required for determination of bromine, chlorine dioxide and ozone in the presence of chlorine | ^{g)} Reagent recovers most insoluble iron oxides without digestion | ^{h)} additionally required for samples with hardness values above 300 mg/l CaCO₃ | ⁱ⁾ high range by dilution | ^{j)} including stirring rod, 10 cm