

204 PtCo

Hazen	
10 - 500 mg/l Pt	
(APHA) Platinum Cobalt Standard Me- thod	

Instrument specific information

The test can be performed on the following devices. In addition, the required cuvette and the absorption range of the photometer are indicated.

Instrument Type	Cuvette	λ	Measuring Range			
MD 100, MD 600, MD 610, MD 640, MultiDirect	ø 24 mm	430 nm	10 - 500 mg/l Pt			
XD 7000, XD 7500	ø 24 mm	455 nm	10 - 50	00 mg/l Pt		
Material						
Required material (partly optional):						
Reagents		Packagin	g Unit	Part Number		
no reagent required						

Application List

- Waste Water Treatment
- Drinking Water Treatment
- Raw Water Treatment

Preperation

1. Sample collection, preservation and storage:

Pour the water sample into clean glass or plastic containers and analyse as soon as possible after the sample is taken. If this is not possible, fill the container right up to the top and seal tightly. Do not stir the sample and avoid lengthy contact with the air. The sample may be stored in a dark place at a temperature of 4°C for 24 hours. Before carrying out any measurements, the water sample should be brought up to room temperature.

Notes

1. This colour scale was originally developed by A. Hazen as a visual comparison scale. It is therefore necessary to ascertain whether the extinction maximum of the water sample is in the range between 420 and 470 nm, as this method is only suitable for water

samples with yellowish to yellowish-brown colouration. Where applicable, a decision should be made based on visual inspection of the water sample. 2. This method is calibrated on the basis of the standards specified by "Standard Methods for the Examination of Water and Wastewater" (also see EN ISO 7887:1994).

Pt-Co colour unit ^= 1 mg/L of platinum as chloroplatinate ion 3. Colour may be expressed as "true" or "apparent" colour. The apparent colour is defined as the colour of a solution due to dissolved substances and suspended particles in the sample.

This manual describes the determination of true colour by filtration of the water sample. To determine the apparent colour, non-filtrated deionised water and sample are measured. 4. The estimated detection limit for this method is 15 mg/L Pt.

Implementation of the provision Colour, true and apparent

Select the method on the device

For this method, no ZERO measurements are to be carried out with the following devices: XD 7000, XD 7500








with a pre-rinsed filter (pore in the blank. size 0.45 µm).

Filter approx. 50 ml sample Put 10 ml deionised water Close vial(s).







Place **blank** in the sample chamber. • Pay attention to the positioning.

Press the ZERO button.

Remove the vial from the sample chamber.



Empty vial.

For devices that require no ZERO measurement, start here.



Fill 24 mm vial with **10 ml** prepared sample .



Place **sample vial** in the sample chamber. • Pay attention to the positioning.



Press the **TEST** (XD: **START**) button.

The result in Pt-Co units appears on the display.

Hazen / 204

Chemical Method

(APHA) Platinum Cobalt Standard Method

Appendix

According to

DIN 7887-C1 (WL 430, 455 nm; Standard: 410 nm)

^{a)} determination of free, combined and total | ^{b)} Reactor is necessary for COD (150 °C), TOC (120 °C) and total -chromium, - phosphate, -nitrogen, (100 °C) | ^a MultiDirect: Adapter is necessary for Vacu-vials[®] (Order code 19 20 75) | ^a) Spectroquant[®] is a Merck KGaA Trademark | ^a) alternative reagent, used instead of DPD No.1/No.3 in case of turbidity in the water sample caused by high concentration of calcium and/or high conductivity | ^a additionally required for determination of bromine, chlorine dioxide and ozone in the presence of chlorine | ^a Reagent recovers most insoluble iron oxides without digestion | ^b additionally required for samples with hardness values above 300 mg/l CaCO₃ | ^a high range by dilution | ^a including stirring rod, 10 cm

Hazen / 204