



REVISIONS			
ZONE	REV	DESCRIPTION	DATE
			APPROVED

IIIG02N004

C



- Notes:
- 1) Sensor Material: Epoxy, Polyurethane, PVC, 316 SS
 - 2) Oxygen measurements should be take 2ft. below the surface and 3ft. from the wall.
 - 3) Measurement element surface should not be facing downward in order to avoid collecting air bubbles.

UNLESS OTHERWISE SPECIFIED, DIMENSIONS ARE IN INCHES. TOLERANCES ON:				CONTRACT NO.				 <div>80 Whisperwood Blvd. Suite 107 Slidell, LA 70458</div>			
FRACTIONS				SIGNATURES				DATES			
DECIMALS				DRAFTSMAN							
ANGLES				CHECKER							
MATERIAL				DESIGNER							
SPEC				PROJ. ENGR.							
FINISH				RELEASED							
-				APPROVAL							
DO SENSOR OUTLINE & MOUNTING								<div> <div>SIZE</div> <div>FSOM NO.</div> <div>DRAWING NO.</div> <div>REV</div> </div> <div> <div>B</div> <div></div> <div>IIIG02N004</div> <div>C</div> </div>			
SCALE NTS								SHEET 1 OF 1			