



The Quick Pool Check®i procedure:

1. Turn the meter on and press Start which begins the 35 second timer. Dip the test strip for 5 seconds without motion into pool water sample. (Sample should be collected from a depth of 18 inches.)
2. Remove test strip and dry excess water from test strip with paper towel. Position strip into Auto Feed Strip Track and Pool Check®i will automatically withdraw track and begin scan when timer reaches zero.
3. Read your results on the Display Screen.

