

Certificate Of Analysis

Product Information:

Part Number	00653-27
Description	OAKTON 2070uS/CM COND STANDARD, 500 ML
Lot Number	RU2
Expiration Date	06/30/2014
Method Of Analysis	CONDUCTIVITY

Actual Analysis:

<u>Acceptance Range</u>	<u>Results</u>	<u>Unit Of Measure</u>
2070 US/CM +/- 10 US/CM AT 25 C	2076	US/CM AT 25 C

Traceability:

<u>N.I.S.T Chemical</u>	<u>SRM#</u>
POTASSIUM CHLORIDE	999b

Certificate Date:06/27/2013

Approved By:



Quality Representative

thermoscientific.com/water

© 2012 Thermo Fisher Scientific Inc. All rights reserved. All trademarks are the property of Thermo Fisher Scientific Inc. & its subsidiaries

Water Analysis Instruments

North America
Toll Free: 1-800-225-1480
Tel: 1-978-232-6000
info.water@thermo.com

244721-001 Rev.H

Netherlands
Tel: (31) 033-2463887
info.water.uk@thermo.com

China
Tel: (86) 21-68654588
wai.asia@thermofisher.com

India
Tel: (91) 22-4157-8800
wai.asia@thermofisher.com

Singapore
Tel: (65) 6778-6876
wai.asia@thermofisher.com

Japan
Tel: (81) 045-453-9175
wai.asia@thermofisher.com

Australia
Tel: (613) 9757-4300
in Australia (1300) 735-295
InfoWaterAU@thermofisher.com

Thermo
SCIENTIFIC

Part of Thermo Fisher Scientific