

## Certificate of Analysis

### Product Information:

<b>Product Code</b>	EC44250BT
<b>Description</b>	50ppm@25 deg C 442 Conductivity Solution
<b>Batch Number</b>	S094/01
<b>Manufacturing Date</b>	28-Feb-2014
<b>Expiration Date</b>	28-Feb-2015

### Actual Analysis

Measurement	Acceptance Range	Results
Conductivity (@25degC)	75 uS/cm +/- 5%	Pass

### Method of Standardization

This solution is standardized using a conductivity meter that is calibrated by NIST traceable standard. A platinized platinum cell was used for measurement.

Approved By



**Chan Cheow Kwang**  
Manager - QA/Regulatory