

Certificate of Analysis

Product Information:

Product Code	ECCON5000BT
Description	5000 uS/cm @ 25 deg C Conductivity Solution
Batch Number	S193/01
Manufacturing Date	10-May-2013
Expiration Date	10-May-2016

Actual Analysis

Measurement	Acceptance Range	Results
Conductivity (@25 deg C)	5000 uS/cm +/- 1%	Pass

Method of Standardization

This solution is standardized using a conductivity meter that is calibrated by a NIST traceable standard. A platinized platinum cell was used for measurement.

Approved By



Chan Cheow Kwang
Manager - QA/Regulatory