

Certificate of Analysis

Product Information:

Product Code	ECCON2764BT
Description	2764 uS/cm @ 25 deg C Conductivity Solution
Batch Number	S333/01
Manufacturing Date	16-Aug-2013
Expiration Date	16-Aug-2016

Actual Analysis

Measurement	Acceptance Range	Results
Conductivity (@25 deg C)	2764 uS/cm +/- 1%	Pass

Method of Standardization

This solution is standardized using a conductivity meter that is calibrated by a NIST traceable standard. A platinized platinum cell was used for measurement.

Approved By



Chan Cheow Kwang
Manager - QA/Regulatory