

## Certificate of Analysis

### Product Information:

<b>Product Code</b>	ECBU686BT
<b>Description</b>	6.86 pH @ 25 deg C pH buffer solution
<b>Batch Number</b>	S253/01
<b>Manufacturing Date</b>	21-Jun-2013
<b>Expiration Date</b>	21-Jun-2016

### Actual Analysis

Measurement	Acceptance Range	Results
pH value (@ 25 deg C)	6.86+/-0.01 pH	Pass

### Method of Standardization

A mV instrument with a sensor calibrated to NIST traceable standards was used for cross checking the solution value.

Approved By



**Chan Cheow Kwang**  
**Manager - QA/Regulatory**