

Certificate of Analysis

Product Information:

Product Code	ECBU1BT
Description	1.68 pH @25 deg C pH Buffer Solution
Batch Number	S273/01
Manufacturing Date	05-Jul-2013
Expiration Date	05-Jul-2016

Actual Analysis

Measurement	Acceptance Range	Results
pH value (@ 25 deg C)	1.68+/-0.01 pH	Pass

Method of Standardization

A mV instrument with a sensor calibrated to NIST traceable standards was used for cross checking the solution value.

Approved By



Chan Cheow Kwang
Manager - QA/Regulatory