

117934

100 Tests

MQuant™

Total Hardness Test

Gesamthärte-Test

mg/l
CaCO₃mmol/l
CaCO₃

<150



<1.5

150 - 250



1.5 - 2.5

>250



>2.5



1 s



1 min



1°d \triangleq 10 mg/l CaO \triangleq 1.25 °e \triangleq 1.78 °f \triangleq 17.8 mg/l CaCO₃
 \triangleq 0.18 mmol/l CaCO₃ (Ca)

