

Certificate of Analysis

Product Information:

Product Code	EC442300BT
Description	300 PPM @ 25 deg C 442 Conductivity solution
Batch Number	S244/01
Manufacturing Date	13-Jun-2014
Expiration Date	13-Jun-2017

Actual Analysis

Measurement	Acceptance Range	Results
Conductivity (@25degC)	449 uS/cm +/- 2%	Pass

Method of Standardization

This solution is standardized using conductivity meter that is calibrated by NIST traceable standard. A platinized platinum cell was used for measurements.

Approved By



Chan Cheow Kwang
Manager - QA/Regulatory