

GFS Chemicals, Inc.
Columbus, Ohio 43223

LOT ANALYSIS

ITEM: 8007 AMCO CLEAR TURBIDITY STANDARD, 1.0 NTU

LOT#:C469970

Test	PASS/FAIL	NUMERICAL RESULT
Turbidity 1.0 NTU	PASS	1.0 NTU
Absorbance at 455nm (100mm pathlength)	PASS	0.0119
Lot Number of Baseline Material	PASS	C469545

TRACEABLE TO N.I.S.T. (Y/N)? Y

Comment:

Reported by: Rachel Stapf

QC Supervisor: Rachel Stapf

Quality Assured to Retest Point: 12 months from shipment

C/A Print Date: 01/28/2015

Not for direct use in food, cosmetics, finished pharmaceuticals or drug products. Supplier is not responsible for compliance with FDA Current Good Manufacturing Practice (CGMP) requirements, including without limitation those for finished drug products in 21 C.F.R Parts 210 and 211. Consult warranty limitations at www.gfschemicals.com/statics/documents/aboutus/termsandconditions.html For resale by GFS authorized distributors only.