GFS Chemicals, Inc. Columbus, Ohio 43223

LOT ANALYSIS ITEM: 8000 AMCO CLEAR TURBIDITY STANDARD, 0.0 NTU

LOT#:C474813

Test	PASS/FAIL	NUMERICAL RESULT
Turbidity (instrument specific) = 0.04 NTU</td <td>PASS</td> <td><0.04 NTU</td>	PASS	<0.04 NTU
Turbidity (HF Scientific instruments) 0.02 NTU	PASS	0.02 NTU
Absorbance at 455 nm against pure methanol < 0.001	PASS	< 0.001
Lot Number of Baseline Material	PASS	C472596

TRACEABLE TO N.I.S.T. (Y/N)? Y

Comment:

Reported by: Stephen Bailey QC Supervisor: Rachel Stapf

Quality Assured to Retest Point: 12 months from shipment

C/A Print Date: 01/28/2015

Not for direct use in food, cosmetics, finished pharmaceuticals or drug products. Supplier is not responsible for compliance with FDA Current Good Manufacturing Practice (CGMP) requirements, including without limitation those for finished drug products in 21 C.F.R Parts 210 and 211. Consult warranty limitations at www.gfschemicals.com/statics/documents/aboutus/termsandconditions.html For resale by GFS authorized distributors only.