

IIG03N010		REVISIONS	
ZONE	REV	DESCRIPTION	DATE
			APPROVED

1/4" TUBE FITTING
InsitelG Supplied

1.90

CONNECTION CABLE
TO ANALYZER

1-1/2" NPT

1.39

5.9

1.00

2.65

Notes:

- 1) Material: Epoxy, Polyurethane, & PVC
- 2) For best accuracy, Suspended Solids measurements should be taken 2ft. below the surface and 3ft. from the wall.

UNLESS OTHERWISE SPECIFIED: DIMENSIONS ARE IN INCHES
TOLERANCES ON:
FRACTIONS DECIMALS ANGLES
MATERIAL Listed
SPEC
FINISH -

CONTRACT NO.	
SIGNATURES	DATES
DRAFTSMAN	
CHECKER	
DESIGNER	
PROJ. ENGR.	
RELEASED	
APPROVAL	

InsitelG		80 Whisperwood Blvd. Suite 107 Sidell, LA 70458	
Outline & Mounting			
Low Range 15L/25L Sensor		DRAWING NO.	IIG03N010
SIZE	FSCM NO.	REV	
B			
SCALE		NTS	
		SHEET 1 OF 1	