

# Certificate Of Analysis

## Product Information:

Part Number	00653-23
Description	OAKTON 23uS/CM COND STANDARD, 500 ML
Lot Number	SU1
Expiration Date	06/30/2015
Method Of Analysis	CONDUCTIVITY

## Actual Analysis:

<u>Acceptance Range</u>	<u>Results</u>	<u>Unit Of Measure</u>
23 US/CM + /- 1 US/CM AT 25 C	23	US/CM AT 25 C

## Traceability:

<u>N.I.S.T Chemical</u>	<u>SRM#</u>
POTASSIUM CHLORIDE	999b

Certificate Date:06/11/2014

Approved By:

Quality Representative

[thermoscientific.com/water](http://thermoscientific.com/water)

© 2012 Thermo Fisher Scientific Inc. All rights reserved. All trademarks are the property of Thermo Fisher Scientific Inc. & its subsidiaries

Water Analysis Instruments

**North America**  
Toll Free: 1-800-225-1480  
Tel: 1-978-232-6000  
info.water@thermo.com

244721-001 Rev.H

**Netherlands**  
Tel: (31) 033-2463887  
info.water.uk@thermo.com

**China**  
Tel: (86) 21-68654588  
wai.asia@thermofisher.com

**India**  
Tel: (91) 22-4157-8800  
wai.asia@thermfisher.com

**Singapore**  
Tel: (65) 6778-6876  
wai.asia@thermfisher.com

**Japan**  
Tel: (81) 045-453-9175  
wai.asia@thermfisher.com

**Australia**  
Tel: (613) 9757-4300  
in Australia (1300) 735-295  
InfoWaterAU@thermfisher.com

**Thermo**  
SCIENTIFIC  
Part of Thermo Fisher Scientific