



REVISIONS			
ZONE	REV	DESCRIPTION	DATE
			APPROVED

11G03N004

A




- Notes:
- 1) Material: Epoxy, Polyurethane, & PVC
 - 2) For best accuracy, Suspended Solids measurements should be taken 2ft. below the surface and 3ft. from the wall.

UNLESS OTHERWISE SPECIFIED: DIMENSIONS ARE IN INCHES. TOLERANCES ON: FRACTIONS DECIMALS ANGLES				CONTRACT NO.				<div>  <div> 1550 W. Lindburg Dr. Stidell, LA 70458 </div> </div>			
MATERIAL —				SIGNATURES		DATES		<div> <div>TSS SENSOR OUTLINE & MOUNTING</div> <div> <div> <div>SIZE</div> <div>FSCM NO.</div> <div>DRAWING NO.</div> </div> <div> <div>11G03N004</div> <div>REV</div> </div> </div> </div>			
				DRAFTSMAN							
				CHECKER							
				DESIGNER							
SPEC —				PROJ. ENGR.							
FINISH —				RELEASED				SHEET 1 OF 1			
				APPROVAL							

REVISIONS			
ZONE	REV	DESCRIPTION	DATE
			APPROVED

IIIG03N005 A



- Notes:
- 1) Material: Epoxy, Polyurethane, & PVC
 - 2) For best accuracy, Suspended Solids measurements should be taken 2ft. below the surface and 3ft. from the wall.

UNLESS OTHERWISE SPECIFIED: DIMENSIONS ARE IN INCHES. TOLERANCES ON: FRACTIONS DECIMALS ANGLES MATERIAL SPEC FINISH			
CONTRACT NO.			
SIGNATURES DATES			
DRAFTSMAN CHECKER DESIGNER PROJ. ENGR. RELEASED APPROVAL			
Insite IG 1550 W. Lindburg Dr. Slidell, LA 70458			
TSS SENSOR MOUNTING			
DRAWING NO. IIIG03N005			
SIZE B FSCM NO. SCALE NTS			
SHEET 1 OF 1			
REV A			