



ALP Co., Ltd.

Scientific & Medical
Manufacturer

3-3-10, Midorigaoka
Hamura-shi
Tokyo 205, Japan
TEL. 81-42-579-0531
FAX. 81-42-579-0533

Jun. 20, 2005

CERTIFICATE OF QUALITY

To whom it may concern,

We hereby certify that the following products were manufactured by us and passed our test and inspection for performance, structure and operation by our quality control division.

Production Date : Jun. 17, 2005
Production No. : 801830
Product Name : Autoclave (Automatic steam sterilizer)
Model KT-2322

Item	Test specifications	Test Result
1. Water Pressure Test	0.36MPa	GOOD
2. Safety Valve Test	0.18MPa	GOOD
3. Operating Test	0.16MPa - 127°C	GOOD
4. Electric Source	A. C. 100V, 1 ϕ , 50/60Hz, 1.4KW	GOOD
5. Insulation Resistance Test	500V. D. C. 100Mohm~	GOOD
6. Voltage Resistance Test	1,500V - 1 minute	GOOD
7. Electric Circuit Check		GOOD
8. Tubing Check		GOOD
9. Appearance		GOOD

Yours faithfully,

ALP Co., Ltd.

Junichi Endo
General Manager