

## Certificate of Analysis

### Product Information:

<b>Product Code</b>	ECCON1118BT
<b>Description</b>	111.8 mS/cm @ 25 deg C Conductivity Solution
<b>Batch Number</b>	S063/01
<b>Manufacturing Date</b>	08-Feb-2013
<b>Expiration Date</b>	08-Feb-2016

### Actual Analysis

Measurement	Acceptance Range	Results
Conductivity (@ 25 deg C)	111.8 mS/cm +/- 1%	Pass

### Method of Standardization

This solution is standardized using a conductivity meter that is calibrated by NIST traceable standard. A platinized platinum cell was used for measurement.

Approved By



**Chan Cheow Kwang**  
**Manager - QA/Regulatory**