

## Certificate of Analysis

### Product Information:

<b>Product Code</b>	EC4423000BT
<b>Description</b>	3000 PPM @ 25 deg C 442 Conductivity solution
<b>Batch Number</b>	S483/01
<b>Manufacturing Date</b>	29-Nov-2013
<b>Expiration Date</b>	29-Nov-2016

### Actual Analysis

Measurement	Acceptance Range	Results
Conductivity (@25degC)	4.02 mS/cm +/- 1%	Pass

### Method of Standardization

This solution is standardized using conductivity meter that is calibrated by NIST traceable standard. A platinized platinum cell was used for measurements.

Approved By



**Chan Cheow Kwang**  
Manager - QA/Regulatory