

Certificate of Analysis

Product Information:

Product Code	ECNA45PPT
Description	NaCl 45 ppt Solution
Batch Number	S024/01
Manufacturing Date	10-Jan-2014
Expiration Date	10-Jan-2017

Actual Analysis

Measurement	Acceptance Range	Results
Conductivity (@25 deg C)	74.12 mS/cm +/- 1 %	Pass

Method of Standardization

This solution is standardized using a conductivity meter that is calibrated by a NIST traceable standard. A platinized platinum cell was used for measurement.

Approved By



Chan Cheow Kwang
Manager - QA/Regulatory