

Certificate of Analysis

Product Information:

Product Code	ECCON1413BT
Description	1413 uS/cm @ 25 deg C Conductivity Solution
Batch Number	S223/01
Manufacturing Date	31-May-2013
Expiration Date	31-May-2016

Actual Analysis

Measurement	Acceptance Range	Results
Conductivity @ 25 deg C	1413 uS/cm +/- 1%	Pass

Method of Standardization

This solution is standardized using a conductivity meter that is calibrated by NIST traceable standard. A platinized platinum cell was used for measurement.

Approved By



Chan Cheow Kwang
Manager - QA/Regulatory